



Morrison Medical Economy Best Strap Instructions for use

ENGLISH

Important!
*Read all instructions thoroughly
before using the
Morrison Best Strap™*

Only qualified and properly trained personnel should use the Morrison Medical Best Strap. Any persons using this strap should practice application and be thoroughly familiar with all medical protocols and aspects of its use **BEFORE** using. These instructions are provided as a guide only. Use protocols and techniques established by your local medical control.

1

Make sure straps are not twisted.

Make certain the strap is not too tight or too loose. You should be able to place four fingers (with your hand flat) between the patient and the strap.

Caution: Begin with ankle straps to assure proper placement of shoulder straps and assure adequate circulation.

1. Make sure the patient is placed correctly on the board.
2. Place the "Economy Best Strap" along the patient's body so that the shoulder straps (Shoulder Strap) lie across both shoulders and the center strap (Anchor) lies along anterior midline. Hook and loop side up. Do not secure shoulder straps at this time.
3. Starting at one side of the ankle strap (Ankle); thread the loop material strap to secure. Repeat for other side of this strap. Always keep Anchor strap centered.
4. Repeat the above process with the remaining body straps; thigh and pelvic and the chest strap.
5. Finish with the two shoulder straps; thread the loop material strap through the handhold in the backboard. Pull firmly and press onto hook material strap to secure. Repeat for other side of this strap. Always keep anchor strap centered.
6. Double-check all straps.

3

Description/Intended Use:

The Morrison Medical Best Strap is a strap system comprised of 4 body and 2 over the shoulder straps with hook and loop fasteners. Designed to secure a patient's body at the shoulder, chest, pelvis, thigh and ankle while laying on a backboard.

Warning:

CHECK PATIENT FREQUENTLY!

IMPROPER USE COULD RESULT IN SERIOUS INJURY OR DEATH.

Inspect strap before each use. Destroy and discard if you observe:

- Broken stitching
- Frayed or cut fabric, straps or closures.
- Other signs of deterioration

Use of products with these conditions could result in product failure resulting in serious injury or death.

2

To Release

Begin with shoulder straps and separate hook from loop on each strap. Continue until all straps are separated.

Cleaning

Secure hook and fasteners to each other. Machine wash, air dry. Do not exceed 108°F. Do not use bleach. Washing inside laundry bag recommended.

Manufactured by:
Morrison Medical
3735 Paragon Drive
Columbus, Ohio 43228-4446



4

